

These instructions serve as guidelines for getting the maximum accuracy potential and avoiding contamination in the use of calibration gas mixtures contained in disposable gas cylinders.

Calibration gas mixtures are subject to contamination from the sampling system, atmospheric moisture, and oxygen. These general instructions will aid in eliminating common operator errors when using the gas mixtures and supplied pressure regulators.

